

Gas and Vapor Detection Instruments for Industrial Health and Safety Applications

680 Fairfield Court, P.O. Box 979, Ann Arbor, MI 48106-0979, Phone: 734-761-1270 Fax: 734-761-3220
www.enmet.com info@enmet.com



**Notice to ENMET Customers With Regards to Carbon Monoxide Sensor Recall of
City Technology 4CF Sensors**

The recent widely published recall of instruments which used the City Technology 4CF carbon monoxide sensors made during the May 2007, to June 2008, period has no impact to ENMET customers. The referenced CO sensor was not then, nor is it now, used in any ENMET portable or stationary gas detectors. As always, instrument users should exhibit due diligence by performing regular field tests and recalibration following schedules recommended by their instrument maker and found in instrument manuals and/or website postings.